

The following guidelines must be considered when working with pressurized equipment within the glove box system:

1. It is important that regular maintenance procedures are carried out on instrumentation operating at higher pressures than the glove box. A sudden component failure could cause damage to the glove box system.
2. Care must be taken in the selection of the most suitable fittings for use with high pressure equipment.